

Certificate of Analysis

No. CA11184

Product: Eichrom's Nickel Resin cartridges
Lot #: NIAR11A

Eichrom's Nickel Resin Lot # NIAR11A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	90.8%
Ave. % Ni in Load/Rinse	<1%	0 %
Ave. % Ni in Strip	>99%	100 %


Jill Bryant
Quality System Coordinator

August 24, 2011