

Certificate of Analysis

No. CA12193

Product: Eichrom's Nickel Resin cartridges


Lot #: NIAR12B

Eichrom's Nickel Resin cartridge Lot # NIAR12B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	92.9 %
Ave. % Ni in Load/Rinse	<1%	0.1 %
Ave. % Ni in Strip	>99%	99.9 %

Jill Bryant
Quality System Coordinator



October 18, 2012