

# Certificate of Analysis

No. CA 15055


Product: Eichrom's Nickel Resin Cartridges  
Lot #: NIAR15A

Eichrom's Nickel Resin cartridge Lot # NIAR15A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	97.7 %
Ave. % Ni in Load/Rinse	<1%	0.4 %
Ave. % Ni in Strip	>99%	99.6 %

Jill Bryant  
Quality System Coordinator



March 26, 2015