

Certificate of Analysis


No. CA08108

Product: Eichrom's Nickel Resin cartridges
Lot #: NIAR8A

Eichrom's Nickel Resin cartridge Lot # NIAR8A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	99.9 %
Ave. % Ni in Load/Rinse	<1%	0.1%
Ave. % Ni in Strip	>99%	99.9%



June 9, 2008

Jill Bryant
Quality System Coordinator