

Certificate of Analysis


No. CA 10091

Product: Eichrom's PB Resin
Lot #: PBS050710

Eichrom's PB Resin Lot # PBS050710, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% Lead in Load + Rinse	< 0.1 %	0.00 %
% Lead in Second Rinse	< 1.0 %	0.00 %
% Lead in Strip	> 99 %	100 %
Gravimetric Yield	80-120 %	89.9 %
Chemical Yield	80-120 %	119 %



May 20, 2010

Jill Bryant
Quality System Coordinator