

Certificate of Analysis

No. CA11041


Product: Eichrom's Pre-Filter Resin cartridges

Lot #: PFLR11A

Eichrom's Pre-Filter Resin cartridges, Lot PFLR11A, has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Flow Rate	> 0.80 ml/min.	Pass



Jill Bryant
Quality System Coordinator

February 21, 2011