

Certificate of Analysis


No. CA11106

Product: Eichrom's Pre-Filter Resin cartridges
Lot #: PFLR11B

Eichrom's Pre-Filter Resin cartridges, Lot PFLR11B, has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Flow Rate	> 0.80 ml/min.	Pass



Jill Bryant
Quality System Coordinator

May 5, 2011