

Certificate of Analysis


No. CA08062

Product: Eichrom's Pre-Filter Resin cartridges
Lot #: PFLR8A

Eichrom's Pre-Filter Resin cartridges, Lot PFLR8A, has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Flow Rate	> 0.80 ml/min.	Pass



March 31, 2008

Jill Bryant
Quality Control Chemist