

# Certificate of Analysis

No. CA 17118

Product: Eichrom's 25mm Resolve™ Filters in Disposable Funnels

Date of Manufacture: 20-Jul-17

Lot #: RFDF071917F

Eichrom's 25mm Resolve™ Filters in Disposable Funnels Lot # RFDF071917F has been manufactured according to Eichrom standard operating procedures. The filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	39.0
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	35.0
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

July 20, 2017