eichrom TECHNOLOGIES
A GCI COMPANY

Certificate of Analysis

No. CA 17202

Product:	Eichrom's 25mm Resolve™ Filters in Disposable Funnels
Lot #:	RFDF122017F
Date of Manufacture:	22-Dec-2017

Eichrom's 25mm Resolve™ Filters in Disposable Funnels Lot # RFDF122017F has been manufactured according to Eichrom standard operating procedures. The filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Am-241 FWHM avg keV (8 replicates)	<50.0	38.6
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	35.3
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0

Sarah McAlister
Quality Manager

22 Dec 17
December 22, 2017

Eichrom Technologies LLC
1955 University Lane, Lisle, IL 60532, USA
www.eichrom.com

npo
eichrom