

**Certificate of Analysis**

No. CA10017


Product: Eichrom's Sr Resin

Lot #: SRA012510

Eichrom's Sr Resin Lot # SRA012510 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	$\geq 100$	144
% Gravimetric Recovery	$\leq 105$	102.5
% Chemical Recovery, Corrected	90-110	95.2
% Sr stripped in 0.05M HNO <sub>3</sub>		
Least acceptable mean (5 rep)	85.0	98.8
Max. standard deviation	2.5	1.0
%Ca stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.25	0.05
Max. standard deviation	0.25	0.00
%Ba stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.50	0.36
Max. standard deviation	0.50	0.09
%Y stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.25	0.0
Max. standard deviation	0.25	0.0
%Y in 8M HNO <sub>3</sub> load and rinse		
Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0
Raw support resin gravity flow rate, mL/min	0.60-0.80	0.72



Joel Williamson  
Quality Manager

Prepared February 5, 2010