

Certificate of Analysis

No. CA07077

Product: Eichrom's Sr Resin

Lot #: SRA05077

Eichrom's Sr Resin, Lot # SRA05077 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	≥ 100	128
Gravimetric Recovery	$\leq 105\%$	102 %
Chemical Recovery, Corrected	90-110 %	98.6 %



June 20, 2007

Jill Bryant
Quality Control Chemist