

## Certificate of Analysis

No. CA06073

Product: Eichrom's Sr Resin

Lot #: SRA06146

Eichrom's Sr Resin, Lot # SRA06146 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	$\geq 100$	142
Gravimetric Recovery	$\leq 105\%$	104 %
Chemical Recovery, Corrected	90-110 %	96.9 %



Jill Bryant

Quality Control Chemist

Prepared: July 11, 2006