

Certificate of Analysis

No. CA07138


Product: Eichrom's Sr Resin

Lot #: SRA07277

Eichrom's Sr Resin Lot # SRA07277 has been manufactured according to Eichrom standard operating procedures. The resin has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	≥ 100	124
Gravimetric Recovery	$\leq 105\%$	104 %
Chemical Recovery, Corrected	90-110 %	93.9 %


Jill Bryant
Quality Control Chemist

Prepared: September 24, 2007