


**Certificate of Analysis**  
No. CA11204

Product: Eichrom's Sr Resin  
Lot #: SRA091011

Eichrom's Sr Resin Lot # SRA091011 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	$\geq 100$	120
% Gravimetric Recovery	$\leq 105$	103.3
% Chemical Recovery, Corrected	90-110	93.6
% Sr stripped in 0.05M HNO <sub>3</sub>		
Least acceptable mean (5 rep)	85.0	98.1
Max. standard deviation	2.5	1.0
%Ca stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.25	0.00
Max. standard deviation	0.25	0.00
%Ba stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.50	0.00
Max. standard deviation	0.50	0.00
%Y stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.25	0.0
Max. standard deviation	0.25	0.0
%Y in 8M HNO <sub>3</sub> load and rinse		
Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0
Raw support resin gravity flow rate, mL/min	0.60-0.80	0.65

  
Sarah Teichma  
Quality Supervisor

9-19-11  
September 19, 2011