

Certificate of Analysis
No. CA11224

Product: Eichrom's Sr Resin
Lot #: SRA100711

Eichrom's Sr Resin Lot # SRA100711 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	≥ 100	131.2
% Gravimetric Recovery	≤ 105	101.2
% Chemical Recovery, Corrected	90-110	92.9
% Sr stripped in 0.05M HNO ₃		
Least acceptable mean (5 rep)	85.0	98.8
Max. standard deviation	2.5	0.7
%Ca stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.25	0.00
Max. standard deviation	0.25	0.00
%Ba stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.50	0.14
Max. standard deviation	0.50	0.01
%Y stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.25	0.0
Max. standard deviation	0.25	0.0
%Y in 8M HNO ₃ load and rinse		
Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0
Raw support resin gravity flow rate, mL/min	0.60-0.80	0.65


Sarah Teichma
Quality Supervisor

10-24-11
October 24, 2011