

## Certificate of Analysis No. CA12039

Product: Eichrom's Sr Resin

Lot #: SRS030612

Eichrom's Sr Resin Lot # SRS030612 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## **ANALYSIS RESULTS**

Test Value	Specification	Result
Appearance Distribution Ratio, Sr in 3.1 N nitric acid % Gravimetric Recovery % Chemical Recovery, Corrected	White beads ≥ 100 ≤105 90-110	Pass 113.3 99.8 97.4
% Sr stripped in 0.05M HNO <sub>3</sub> Least acceptable mean (5 rep) Max. standard deviation	85.0 2.5	95.9 2.5
%Ca stripped in 0.05M HNO <sub>3</sub> Maximum acceptable mean (5 rep) Max. standard deviation	0.15 0.05	0.03 0.05
%Ba stripped in 0.05M HNO <sub>3</sub> Maximum acceptable mean (5 rep) Max. standard deviation	0.25 0.10	0.00 0.00
%Y stripped in 0.05M HNO <sub>3</sub> Maximum acceptable mean (5 rep)  Max. standard deviation	0.15 0.05	0.00 0.00
%Y in 8M HNO <sub>3</sub> load and rinse Least acceptable mean (5 rep) Max. standard deviation	99.0 2.5	100.0 0.0

Sarah Tejchma Quality Supervisor 3-13-12

March 13, 2012