

## Certificate of Analysis No. CA10249

Product: Eichrom's Sr Resin Lot #: SRS120110

Eichrom's Sr Resin Lot # SRS120110 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## **ANALYSIS RESULTS**

Test Value	<u>Specification</u>	Result
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	≥ 100	122
% Gravimetric Recovery	≤105	100.8
% Chemical Recovery, Corrected	90-110	94.3
% Sr stripped in 0.05M HNO <sub>3</sub>		
Least acceptable mean (5 rep)	85.0	95.9
Max. standard deviation	2.5	1.9
%Ca stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.00
%Ba stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.25	0.00
Max. standard deviation	0.10	0.00
%Y stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.00
0/V in 9M UNO load and since		
%Y in 8M HNO <sub>3</sub> load and rinse Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0
itian. Standard doylanon	2.5	0.0

Joel Williamson Quality Manager December 9, 2010