

Certificate of Analysis

No. CA07158

Product: Eichrom's Sr Resin cartridges

Lot #: SRSR7E

Eichrom's Sr Resin cartridge Lot # SRSR7E has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	≥ 100	136
Gravimetric Recovery	$\leq 105 \%$	103 %
Chemical Recovery, Corrected	90-110 %	95.3 %



November 13, 2007

Jill Bryant
Quality Control Chemist