

**Certificate of Analysis**  
 No. CA11219

Product: Eichrom's TEVA Resin  
 Lot #: TEA092811

Eichrom's TEVA Resin Lot # TEA092811 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.0 N nitric acid	>200	396
% <sup>230</sup> Th stripped in 6M HCl		
Least acceptable mean (4 rep)	93.0	99.9
Max. standard deviation	2.5	0.0
% <sup>230</sup> Th stripped in 0.02M HCl + 0.02M HF		
Maximum acceptable mean (4 rep)	2.0	0.1
Max. standard deviation	1.5	0.0
% <sup>239</sup> Pu stripped in 0.02M HCl + 0.02M HF		
Least acceptable mean (4 rep)	90.0	98.7
Max. standard deviation	2.5	1.1
% <sup>239</sup> Pu stripped in 6M HCl		
Maximum acceptable mean (4 rep)	1.0	0.0
Max. standard deviation	0.5	0.1
Extractant distribution ratio, <sup>230</sup> Th in 3M HNO <sub>3</sub>	>210	302
Extractant distribution ratio, <sup>230</sup> Th in 6M HCl	<0.01	0.002
Raw support resin flow rate, mL/min	0.60-0.80	0.65

Jill Bryant  
 Quality System Coordinator

*Jill Bryant*

October 17, 2011