

Certificate of Analysis
 No. CA12211


Product: Eichrom's TEVA Resin
 Lot #: TEA102912

Eichrom's TEVA Resin Lot # TEA102912 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.0 N nitric acid	>200	416
% ²³⁰ Th stripped in 6M HCl		
Least acceptable mean (4 rep)	93.0	97.7
Max. standard deviation	2.5	0.7
% ²³⁰ Th stripped in 0.02M HCl + 0.02M HF		
Maximum acceptable mean (4 rep)	2.0	0.9
Max. standard deviation	1.5	0.4
% ²³⁹ Pu stripped in 0.02M HCl + 0.02M HF		
Least acceptable mean (4 rep)	90.0	97.3
Max. standard deviation	2.5	0.8
% ²³⁹ Pu stripped in 6M HCl		
Maximum acceptable mean (4 rep)	1.0	0.3
Max. standard deviation	0.5	0.0
Extractant distribution ratio, ²³⁰ Th in 3M HNO ₃	>210	369
Extractant distribution ratio, ²³⁰ Th in 6M HCl	<0.01	0.001
Raw support resin flow rate, mL/min	0.60-0.80	0.70

Jill Bryant
 Quality System Coordinator



November 9, 2012