

Certificate of Analysis

No. CA08270

Product: Eichrom's TEVA Resin


Lot #: TEA12168


Eichrom's TEVA Resin Lot # TEA12168 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Appearance	White beads	Pass
Distribution Ratio, 3.0 N nitric acid	>200	364
% ^{230}Th stripped in 6M HCl		
Least acceptable mean (5 rep)	93.0	99.6
Max. standard deviation	2.5	0.7
% ^{230}Th stripped in 0.02M HCl + 0.02M HF		
Maximum acceptable mean (5 rep)	2.0	0.4
Max. standard deviation	1.5	0.5
% ^{239}Pu stripped in 0.02M HCl + 0.02M HF		
Least acceptable mean (5 rep)	90.0	99.7
Max. standard deviation	2.5	0.1
% ^{239}Pu stripped in 6M HCl		
Maximum acceptable mean (5 rep)	1.0	0.0
Max. standard deviation	0.5	0.0
Raw support resin flow rate, mL/min	0.60-0.80	0.68

TEVA Resin Elution Curve




 Joe Williamson
 Quality Manager

Prepared: January 5, 2009