

Certificate of Analysis
 No. CA10238


Product: Eichrom's TEVA Resin cartridges
 Lot #: TESR10Q

Eichrom's TEVA Resin cartridge Lot # TESR10Q has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.0 N nitric acid	>200	429
% ²³⁰ Th stripped in 6M HCl		
Least acceptable mean (4 rep)	93.0	99.9
Max. standard deviation	2.5	0.0
% ²³⁰ Th stripped in 0.02M HCl + 0.02M HF		
Maximum acceptable mean (4 rep)	2.0	0.2
Max. standard deviation	1.5	0.1
% ²³⁹ Pu stripped in 0.02M HCl + 0.02M HF		
Least acceptable mean (4 rep)	90.0	99.2
Max. standard deviation	2.5	0.1
% ²³⁹ Pu stripped in 6M HCl		
Maximum acceptable mean (4 rep)	1.0	0.1
Max. standard deviation	0.5	0.1
Extractant distribution ratio, ²³⁰ Th in 3M HNO ₃	>210	302
Extractant distribution ratio, ²³⁰ Th in 6M HCl	<0.01	0.002

Jill Bryant
 Quality System Coordinator



November 11, 2010