

Certificate of Analysis
No. CA12070


Product: Eichrom's TEVA Resin cartridges
Lot #: TESR12D

Eichrom's TEVA Resin cartridge Lot # TESR12D has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.0 N nitric acid	>200	483
% ²³⁰ Th stripped in 6M HCl		
Least acceptable mean (4 rep)	93.0	99.5
Max. standard deviation	2.5	0.1
% ²³⁰ Th stripped in 0.02M HCl + 0.02M HF		
Maximum acceptable mean (4 rep)	2.0	0.3
Max. standard deviation	1.5	0.1
% ²³⁹ Pu stripped in 0.02M HCl + 0.02M HF		
Least acceptable mean (4 rep)	90.0	97.1
Max. standard deviation	2.5	0.5
% ²³⁹ Pu stripped in 6M HCl		
Maximum acceptable mean (4 rep)	1.0	0.2
Max. standard deviation	0.5	0.0
Extractant distribution ratio, ²³⁰ Th in 3M HNO ₃	>250	267
Extractant distribution ratio, ²³⁰ Th in 6M HCl	<0.01	0.002

Jill Bryant
Quality System Coordinator



April 13, 2012