

Certificate of Analysis

No. CA11200

Product: Eichrom's TRU Resin

Lot #: TRF090811

Eichrom's TRU Resin Lot # TRF090811 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Eu in 3.1 N nitric acid	≥ 60	102.0
% ^{241}Am stripped in 4M HCl		
Least acceptable mean (4 rep)	95.0	100.1
Max. standard deviation	2.5	0.1
% ^{241}Am stripped in 0.1M ammonium bioxalate		
Maximum acceptable mean (4 rep)	1.0	0.0
Max. standard deviation	0.5	0.1
% ^{239}Pu stripped in 0.1M ammonium bioxalate		
Least acceptable mean (4 rep)	95.0	97.5
Max. standard deviation	2.5	0.3
% ^{239}Pu stripped in 4M HCl		
Maximum acceptable mean (4 rep)	1.0	0.42
Max. standard deviation	0.5	0.1
Extractant distribution ratio, ^{241}Am strip in 0.05M HNO_3	<0.50	0.24



Sarah Teichma
Quality Supervisor

9-15-11

September 15, 2011