


**Certificate of Analysis**  
No. CA12006

Product: Eichrom's TRU Resin  
Lot #: TRS011312

Eichrom's TRU Resin Lot # TRS011312 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Eu in 3.1 N nitric acid	≥ 60	93.5
% <sup>241</sup> Am stripped in 4M HCl		
Least acceptable mean (4 rep)	95.0	99.5
Max. standard deviation	2.5	0.1
% <sup>241</sup> Am stripped in 0.1M ammonium bioxalate		
Maximum acceptable mean (4 rep)	1.0	0.4
Max. standard deviation	0.5	0.0
% <sup>239</sup> Pu stripped in 0.1M ammonium bioxalate		
Least acceptable mean (4 rep)	95.0	97.6
Max. standard deviation	2.5	0.2
% <sup>239</sup> Pu stripped in 4M HCl		
Maximum acceptable mean (4 rep)	1.0	0.2
Max. standard deviation	0.5	0.1
Extractant distribution ratio, <sup>241</sup> Am strip in 0.05M HNO <sub>3</sub>	<0.50	0.24

  
Sarah Teichma  
Quality Supervisor

1-18-12  
January 18, 2012