


Certificate of Analysis
 No. CA10244

Product: Eichrom's TRU Resin
 Lot #: TRS111010

Eichrom's TRU Resin Lot # TRS111010 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|--|----------------------|---------------|
| Appearance | White beads | Pass |
| Distribution Ratio, Eu in 3.1 N nitric acid | ≥ 60 | 92 |
| % ²⁴¹ Am stripped in 4M HCl | | |
| Least acceptable mean (4 rep) | 95.0 | 98.3 |
| Max. standard deviation | 2.5 | 0.4 |
| % ²⁴¹ Am stripped in 0.1M ammonium bioxalate | | |
| Maximum acceptable mean (4 rep) | 1.0 | 0.9 |
| Max. standard deviation | 0.5 | 0.3 |
| % ²³⁹ Pu stripped in 0.1M ammonium bioxalate | | |
| Least acceptable mean (4 rep) | 95.0 | 97.7 |
| Max. standard deviation | 2.5 | 0.2 |
| % ²³⁹ Pu stripped in 4M HCl | | |
| Maximum acceptable mean (4 rep) | 1.0 | 0.2 |
| Max. standard deviation | 0.5 | 0.2 |
| Extractant distribution ratio, ²⁴¹ Am strip in 0.05M HNO ₃ | <0.50 | 0.27 |


 Joel Williamson
 Quality Manager

November 24, 2010