

Certificate of Analysis


No. CA10237

Product: Eichrom's TRU Resin cartridges
Lot #: TRSR10K

Eichrom's TRU Resin cartridge Lot # TRSR10K has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Eu in 3.1 N nitric acid	≥ 60	92
% ^{241}Am stripped in 4M HCl		
Least acceptable mean (4 rep)	95.0	98.3
Max. standard deviation	2.5	0.4
% ^{241}Am stripped in 0.1M ammonium bioxalate		
Maximum acceptable mean (4 rep)	1.0	0.9
Max. standard deviation	0.5	0.3
% ^{239}Pu stripped in 0.1M ammonium bioxalate		
Least acceptable mean (4 rep)	95.0	97.7
Max. standard deviation	2.5	0.2
% ^{239}Pu stripped in 4M HCl		
Maximum acceptable mean (4 rep)	1.0	0.2
Max. standard deviation	0.5	0.2
Extractant distribution ratio, ^{241}Am strip in 0.05M HNO_3	<0.50	0.27



November 11, 2010

Jill Bryant
Quality System Coordinator