

Certificate of Analysis


No. CA11141

Product: Eichrom's TRU Resin cartridge
 Lot #: TRSR11E

Eichrom's TRU Resin cartridge Lot # TRSR11E has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

| <u>Test Value</u> | <u>Specification</u> | <u>Result</u> |
|----------------------------------------------------------------------------------|----------------------|---------------|
| Appearance | White beads | Pass |
| Distribution Ratio, Eu in 3.1 N nitric acid | ≥ 60 | 93.4 |
| % ²⁴¹ Am stripped in 4M HCl | | |
| Least acceptable mean (4 rep) | 95.0 | 99.6 |
| Max. standard deviation | 2.5 | 0.3 |
| % ²⁴¹ Am stripped in 0.1M ammonium bioxalate | | |
| Maximum acceptable mean (4 rep) | 1.0 | 0.3 |
| Max. standard deviation | 0.5 | 0.1 |
| % ²³⁹ Pu stripped in 0.1M ammonium bioxalate | | |
| Least acceptable mean (4 rep) | 95.0 | 97.0 |
| Max. standard deviation | 2.5 | 0.6 |
| % ²³⁹ Pu stripped in 4M HCl | | |
| Maximum acceptable mean (4 rep) | 1.0 | 0.0 |
| Max. standard deviation | 0.5 | 0.0 |
| Extractant distribution ratio, ²⁴¹ Am strip in 0.05M HNO ₃ | <0.50 | 0.24 |



Safah Tejbama
 QC Supervisor

6-20-11

June 20, 2011