

Certificate of Analysis

No. CA06037

Product: Eichrom's UTEVA Resin


Lot #: UTA03136

Eichrom's UTEVA Resin, Lot # UTA03136, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	> 100	230

Signature



Prepared: April 7, 2006

Jill Bryant
Quality Control Chemist