



Certificate of Analysis  
No. CA07122

a GCI Company

Product: Eichrom's UTEVA<sup>®</sup> Resin  
Lot #: UTA07247

Eichrom's UTEVA<sup>®</sup> Resin Lot # UTA07247 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	>100	148
% <sup>230</sup> Th stripped in 5M HCl/0.05M oxalic acid		
Least acceptable mean (4 rep)	90.0	100.8
Max. standard deviation	10.0	1.3
% <sup>230</sup> Th stripped in 1M HCl		
Maximum acceptable mean (5 rep)	0.5	0.1
Max. standard deviation	1.0	0.0
% <sup>233</sup> U stripped in 1M HCl		
Least acceptable mean (5 rep)	90.0	99.0
Max. standard deviation	10.0	0.7
% <sup>233</sup> U stripped in 5M HCl/0.05M oxalic acid		
Maximum acceptable mean (5 rep)	1.0	0.0
Max. standard deviation	1.0	0.0
Raw support resin flow rate, mL/min	0.60-0.80	0.78

  
Joel Williamson  
Quality Manager

Prepared: August 27, 2007