

Certificate of Analysis

No. CA07001

Product: Eichrom's UTEVA ® Resin

Lot #: UTA12156

Eichrom's UTEVA Resin, Lot # UTA12156, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	> 100	161

Signature



Prepared: January 23, 2007

Jill Bryant
Quality Control Chemist