

Certificate of Analysis
 No. CA11109


Product: Eichrom's UTEVA[®] Resin cartridges
 Lot #: UTSR11E

Eichrom's UTEVA[®] Resin cartridges Lot # UTSR11E has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	>100	223
% ²³⁰ Th stripped in 5M HCl/0.05M oxalic acid		
Least acceptable mean (4 rep)	90.0	97.8
Max. standard deviation	10.0	5.6
% ²³⁰ Th stripped in 1M HCl		
Maximum acceptable mean (4 rep)	0.5	0.1
Max. standard deviation	1.0	0.0
% ²³³ U stripped in 1M HCl		
Least acceptable mean (4 rep)	90.0	99.9
Max. standard deviation	10.0	3.0
% ²³³ U stripped in 5M HCl/0.05M oxalic acid		
Maximum acceptable mean (4 rep)	1.0	0.22
Max. standard deviation	1.0	0.1

Jill Bryant
 Quality System Coordinator



May 5, 2011