



Certificate of Analysis

No. CA18044

Product: Eichrom's MnO₂ Resin Cartridges
Lot #: MNAR18A
Date of Manufacture: 27-MAR-18

Eichrom's MnO₂ Resin cartridge Lot # MNAR18A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, MNA062217, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (4 reps)	≥ 95%	98.7 %
Ba-133 in Load solution (4 reps)	<1.0%	0.89 %

Jill Bryant
Quality System Coordinator

March 27, 2018