

Certificate of Analysis
No. CA 18013

Product: Eichrom's Sr Resin
 Lot #: SRA020518
 Date of Manufacture: 6-FEB-2018

Eichrom's Sr Resin Lot # SRA020518 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	≥ 100	115.1
% Sr stripped in 0.05M HNO ₃		
Least acceptable mean (5 rep)	85.0	96.6
Max. standard deviation	2.5	0.6
%Ca stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.25	0.12
Max. standard deviation	0.25	0.03
%Ba stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.50	0.15
Max. standard deviation	0.50	0.00
%Y stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.25	0.0
Max. standard deviation	0.25	0.0
%Y in 8M HNO ₃ load and rinse		
Least acceptable mean (5 rep)	99.0	99.9
Max. standard deviation	2.5	0.0


 Sarah McAlister
 Quality Manager

16 Feb 18
 February 16, 2018