




Certificate of Analysis
No. CA18055

Product: Eichrom's DGA-Normal Resin
Lot #: DNS041818
Date of Manufacture: 18-APR-2018

Eichrom's DGA-Normal Resin Lot # DNS041818 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% ²⁴¹ Am stripped in 0.1M HCl		
Least acceptable mean (4 rep)	97.0	99.4
Max. standard deviation	2.5	0.2
% ²⁴¹ Am stripped in 0.1M HNO ₃		
Maximum acceptable mean (4 rep)	1.0	0.0
Max. standard deviation	0.5	0.3
% ²³³ U stripped in 0.1M HNO ₃		
Least acceptable mean (4 rep)	97.0	98.5
Max. standard deviation	2.5	0.1
% ²³³ U stripped in 0.1M HCl		
Maximum acceptable mean (4 rep)	1.0	0.7
Max. standard deviation	0.5	0.1



Jill Bryant
Quality System Coordinator

April 20, 2018