

Certificate of Analysis


No. CA 18045

Product: Eichrom's Nickel Resin 5mL Columns
Lot #: NI18A
Date of Manufacture: 28-MAR-2018

Eichrom's Nickel Resin Lot # NI18A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, NIA112117, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	99.3 %
Ave. % Ni in Load/Rinse	<1%	0.3 %
Ave. % Ni in Strip	>99%	99.7 %



Jill Bryant
Quality System Coordinator

March 30, 2018