

Certificate of Analysis

No. CA 20179

Product: Eichrom's Nickel Resin Cartridges
Lot #: NIAR20B
Date of Manufacture: 7-OCT-2020

Eichrom's Nickel Resin cartridge Lot NIAR20B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, # NIA112519, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	102.5 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %


Jill Bryant
Quality System Coordinator

October 8, 2020