

Certificate of Analysis

No. CA20197

Product: Eichrom's MnO₂ Resin Cartridges
Lot #: MNAR20G
Date of Manufacture: 4-NOV-2020

Eichrom's MnO₂ Resin cartridge Lot MNAR20G has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, # MNA042320, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (mean, 4 reps)	> 95%	101.5 %
Ba-133 in Load solution (mean, 4 reps)	<1.0%	0.0 %


Jill Bryant
Quality System Coordinator

November 6, 2020