

# Certificate of Analysis

No. CA 20215

Product:Eichrom's: Actinide Resin  
Lot #: ACA113020  
Date of Manufacture: 30-NOV-2020

Eichrom's Actinide Resin Lot # ACA113020 has been manufactured according to Eichrom standard operating procedures. This resin lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.5 M HCl acid	$\geq 300$	540.3

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

December 4, 2020