

Certificate of Analysis
No. CA 20220

Product: Eichrom's Cation Exchange Resin; 50Wx4, 100-200 mesh, Hydrogen Form Cartridges


Lot: C4MR20B

Date of Manufacture: 11-DEC-2020

Eichrom's Cation Exchange resin cartridges, Lot C4MR20B has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, # 060718C4M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Extractables	< 1 mg(residue) / g(resin)	0.72 mg/g
pH	> 4.0	pass
Wet Volume Capacity	min. 1.1 meq/mL	2.1 meq/mL


Jill Bryant
Quality System Coordinator

December 11, 2020