

# Certificate of Analysis

No. CA 21089

Product: Eichrom's 25mm Resolve™ Filters in Disposable Funnels  
Lot #: RFDF030821F  
Date of Manufacture: 8-MAR-2021

Eichrom's 25mm Resolve™ Filters in Disposable Funnels Lot # RFDF030821F has been manufactured according to Eichrom standard operating procedures. The filter membrane used to manufacture this lot, lot 2187, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	37.2
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	36.3
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0

  
Jill Bryant  
Quality System Coordinator

March 25, 2021