

# Certificate of Analysis

No. CA 17200

Product: Eichrom's Cation Exchange Resin; 50Wx8, 200-400 mesh, H<sup>+</sup> Form

Date of Manufacture: 15-Sep-2017

Lot #: 091517C8F

Eichrom's Cation Exchange resin, Lot 091517C8F, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

Test Value	Specification	Result
Extractables	< 1 mg(residue) / g(resin)	0.30 mg/g
pH	<u>≥</u> 4.0	pass
Wet Volume Capacity	min. 1.7 meq/mL	1.9 meq/mL

  
Jill Bryant  
Quality System Coordinator

April 26, 2021