

# Certificate of Analysis


No. CA 21122

Product: Eichrom's Nickel Resin Cartridges  
Lot #: NIAR21B  
Date of Manufacture: 19-MAY-2021

Eichrom's Nickel Resin cartridge Lot NIAR21B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, # NIA031021, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	100.0 %
Ave. % Ni in Load/Rinse	<1%	0.5 %
Ave. % Ni in Strip	>99%	99.5 %

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

May 19, 2021