

# Certificate of Analysis


No. CA21206

Product: Eichrom's LN-2 Resin  
Lot #: L2F081721  
Date of Manufacture: 17-AUG-2021

Eichrom's LN-2 Resin Lot # L2F081721 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White Beads	Pass
Lu Peak Maximum	11 - 15 BV	12.9 BV
Yb Peak Maximum	6.0 - 9.0 BV	7.9 BV
Seperation Factor	$\geq 1.6$	1.7

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

August 25, 2021