

Certificate of Analysis

No. CA21180

Product: Eichrom's LN Resin Cartridges

Lot #: LNSR21C

Date of Manufacture: 20-AUG-2021

Eichrom's LN Resin cartridge Lot LNSR21C has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, LNS051121, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.1 N nitric acid	>700	2528



Sarah McAlister
Quality Manager

23 Aug 2021
August 23, 2021