

## Certificate of Analysis


No. CA 21209

Product: Eichrom's Sr Resin Cartridges  
Lot #: SRSR21F  
Date of Manufacture: 26-AUG-2021

Eichrom's Sr Resin cartridge Lot SRSR21F has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, SRS071621, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

### ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	$\geq 100$	118.5
% Sr stripped in 0.05M HNO <sub>3</sub>		
Least acceptable mean (5 rep)	85.0	93.6
Max. standard deviation	2.5	1.1
%Ca stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.15	0.01
Max. standard deviation	0.05	0.02
%Ba stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.75	0.27
Max. standard deviation	0.10	0.07
%Y stripped in 0.05M HNO <sub>3</sub>		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.00
%Y in 8M HNO <sub>3</sub> load and rinse		
Least acceptable mean (5 rep)	99.0	99.9
Max. standard deviation	2.5	0.0

  
Jill Bryant  
Quality System Coordinator

August 27, 2021