

Certificate of Analysis

No. CA 21220

Product: Eichrom's 25mm Resolve™ Filters in Disposable Funnels
Lot #: RFDF083021F
Date of Manufacture: 30-AUG-2021

Eichrom's 25mm Resolve™ Filters in Disposable Funnels Lot # RFDF083021F has been manufactured according to Eichrom standard operating procedures. The filter membrane used to manufacture this lot, lot 2186, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	38.6
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	35.3
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0


Jill Bryant
Quality System Coordinator

September 22, 2021