

Certificate of Analysis

No. CA 21192


Product: SCX Cartridges, QML
Part No: SCXQML-R100
Lot No: SCXR21C
Manufactured Date: 5-AUG-2021

Eichrom's SCX Cartridge Lot # SCXR21C has been manufactured according to Eichrom's standard operating procedures. The silica used to manufacture this lot, SCX070521, has been tested in accordance with Eichrom's quality assurance procedures and purity testing. This lot of product meets all Eichrom quality specifications.

No products of animal or human origin are used in the manufacture of this product.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Fe Content (μg per cartridge)	<0.6	PASS
Recovery of Ga (%)	>98.0	99.8



Jill Bryant
Quality System Coordinator

August 9, 2021