

Certificate of Analysis

No. CA 21299

Product: Eichrom's Anion Exchange Resin Cartridges

Lot #: A8MR21G

Date of Manufacture: 3-DEC-2021

Eichrom's Anion Exchange resin cartridges Lot A8MR21G have been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 090721A8M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	n.d.
Wet Volume Capacity	≥ 1.2 meq/mL	1.3



Jill Bryant
Quality System Coordinator

December 3, 2021